

The technical proof for clearer,
fuller and richer sound
with Beltone Amaze™

Jennifer Groth, MA

ABSTRACT

Sound quality is a key driver of consumer satisfaction with hearing aids. With a dramatically enriched new platform, Beltone Amaze adds industry-leading capabilities in input dynamic range and frequency bandwidth to its unique sound processing strategy to enable fuller, clearer and richer sound. This paper presents details on how this new platform enhances the performance of Beltone Amaze technically. A companion paper¹ presents perceptual results from listeners who compared Beltone Amaze with other premium hearing aids.

The benefits of amplification are traditionally defined in terms of better audibility for speech and environmental sounds. However, sound quality is a key driver of overall consumer satisfaction with hearing aids and is high on the wish lists of both hearing aid users and Hearing Care Professionals (HCPs). Approximately 40% of persons with moderate-to-severe loss would be motivated to purchase new hearing aids given better sound quality². And according to Beltone market research, 452 independent HCPs in the US, France and Germany named sound quality to be the top attribute to consider in recommending hearing aids to users. Thus, it is important to focus on sound quality in hearing aid development in addition to other benefits. This is a complicated matter, as there is no consensus on what sound quality actually means or how to evaluate it. Regardless, it is certain that sound quality is a subjective experience that is perceived individually, and that it is considered of great importance.

At Beltone, our development is guided by emulating nature's time-tested patterns and strategies. This approach inherently leads us in the direction of "naturalness". In technical terms, sound reproduction that is natural would be equivalent to the sound being transparent acoustically. This means that the sound can be evaluated objectively by comparing its fidelity to the original sound. However, hearing aids deliberately change the sound relative to the original for hearing loss compensation. Thus, fidelity alone cannot be the sole criterion for sound quality. Decisions are needed to determine the optimum degree of transparency. This balance is dependent both on the listening environment as well as the individual. Naturalness can also

refer to behavior, and this is also carefully considered in our hearing aid designs. For example, when in a conversation with others, it is natural to be able to follow the conversational turn-taking and to shift attention at will. Our goal is for hearing aids to allow the sound to be perceived as natural, but also for the wearer to be able to behave in a natural way in their daily listening environments.

The Beltone Amaze takes sound quality to a new level. The processor core is a complete re-design from previous processor generations and takes an industry-leading position. It has been dramatically enriched to maximize processing efficiency and audio quality. To take full advantage of the advanced processing capabilities, all sound processing algorithms have been rewritten for efficiency, quality, and maximum benefit. Sounds are fuller, clearer and richer than ever before. This paper focuses on how technical properties and advances enable the perception of excellent sound quality with the Beltone Amaze. Perceptual evaluation of sound quality demonstrating the preferred sound quality of Beltone Amaze is considered in a companion paper¹.

A new platform

No one feature or characteristic of a hearing instrument can ensure satisfactory sound quality. Mechanical and electroacoustic designs are critical components of how the final product performs and sounds. For today's digital hearing aids, sampling rate, A/D conversion, and processing speed are some additional factors that can affect the sound output of the hearing instrument, either directly or by means of what sound processing they enable. The characteristics and advantages of the advanced platform for Beltone Amaze are summarized in Table 1. In addition to these improvements, doubling the memory capabilities of the platform has enabled additional features in the Beltone Amaze.

Characteristic	What it means	Advantage for Beltone Amaze
A/D conversion (bits)	Defines the range of sound levels that can be digitized with no distortion	Makes full use of the capability of the microphones and enables the industry's highest input dynamic range of 116 dB SPL
Resolution (bits)	Refers to the length of the digital "words" describing the signal; better resolution provides a more accurate reproduction of the sound	Best-in-class audio resolution on par with pro-audio applications, supporting the best sound quality
Processor speed (MHz)	Coordinates arithmetic operations and data transfer	Double the speed of previous Beltone technology provides maximum efficiency for running sound processing and other hearing aid functions
Configurable sampling rate	How often the incoming signal is sampled per unit of time	Allows extended high frequency bandwidth with clean representation of high frequencies
Dual core processing with new radio	Dedicated processors handle wireless functionality and sound processing separately	New radio gives up to 5 dB better sensitivity for better ear-to-ear performance
Power efficiency	20% better efficiency than previous Beltone platform	Enables use time per charge of two full days for rechargeable solution

Table 1. Overview of the Beltone Amaze platform characteristics.

Digital hearing aids have been around for more than two decades. Initially, the label "digital" was used as a promotional term to market hearing aids featuring digital sound processing, and the implication was that sound quality would be vastly improved. Ironically, the digital technology of those first devices imposed limitations that in some ways actually made them inferior to many programmable analog hearing aids of the time. An example of such a limitation was the maximum sound levels that could be digitized without distortion. Typically, sounds with levels nearing 100 dB SPL or greater would be clipped after entering the microphones. This would add a great deal of distortion to the signal that was processed by the amplifier, and resulted in significantly degraded sound quality. As digital technology in hearing aids advanced, such issues were mitigated, although not completely solved.

Input dynamic range is one limitation that is solved in Beltone Amaze. For the first time, the full range of sound that is made available by the Micro-electric-mechanical systems (MEMS) microphones can be utilized, resulting in a clean signal passed to the amplifier. The hearing aid grade MEMS microphones can transduce very high sound levels and are small, highly consistent in performance, and resistant to environmental conditions. These properties make them ideal for use in hearing aids with advanced directional processing. Consistent and stable microphone performance is essential for continued benefit of the unique CrossLink Directionality 2 strategy³ over the lifetime of the hearing aids.

One type of sound for which the input dynamic range of the hearing aids makes an enormous impact on sound

quality is music. While the primary goal of hearing aid design and use is to hear and communicate with others, the hearing aid's ability to faithfully reproduce music is the ultimate test of its sound quality. While there are other acoustic differences between speech and music, what is significant relative to the input range of the hearing aids is that music has a higher crest factor* than speech and is often more intense⁴. Adding to this is the tendency for loudness levels for music to be lower than for speech; that is, a level which is judged as "loud" for speech will be judged less loud for music⁵. The implication is that people will typically listen to music at higher levels, with increased risk of exceeding the input dynamic range of the hearing aids. Figure 1 illustrates how a system with a reduced input dynamic range can clip the peaks of a signal. In this example, violin music was used. It is obvious visually how the signal components above a certain level are cut off. The audible difference is that the music recorded through the system with the reduced input dynamic range sounds "fuzzy" relative to the one with the higher range that does not clip.




Figure 1. A waveform of classic violin music played through a hearing aid with a high input dynamic range (upper panel) and a reduced input dynamic range (lower panel). The clipping of the signal peaks in the waveform of the lower panel result in a great deal of added distortion that cannot be eliminated by other hearing aid processing.

* Crest factor is the difference between the peak value and overall rms value of a signal. It is an indication of how extreme the peaks are in a signal.

Extended high frequency response

One of the determinants of sound quality in any sound reproduction system is the frequency response. The frequency response encompasses both the range of frequencies that are reproduced as well as how smooth or peaky the response is. In terms of smoothness, a goal of the electroacoustic design of hearing aids is to achieve a frequency response that results in flat insertion gain across the range of amplified frequencies. The range itself is also part of this design. While the selection of components, design of the mechanics, and calibration of the system determine the final usable frequency range and the shape of its response, it is the digital platform that sets the theoretical limit for how wide the range can be. Because the sampling rate of the sound can be increased by more than 50% with the new platform, it is possible to extend the frequency range of Beltone Amaze relative to previous Beltone hearing aids. With the selected configuration of sampling rate, a bandwidth extending to 9.5kHz is made possible.

Hearing aids have traditionally been limited in their ability to reproduce high frequencies. However, evidence has built to suggest that frequencies above this traditional bandwidth limitation are more important to speech perception and sound quality than previously assumed. Therefore, it is considered desirable for hearing aids to be able to amplify at higher frequencies. Contributions of high frequencies beyond the typical bandwidth up to 7 kHz include speech, voice and overall quality^{6,7}, source localization⁸, speech intelligibility⁹, identification of talker¹⁰, and word learning in children¹¹. Apart from increased audibility of higher frequency sounds that is possible with extended high frequency bandwidth, all other features are also impacted by the availability of additional high frequency information. Some, such as Personal Sound ID, Sound Shifter, and streaming capabilities are of particular interest.

Impact on Personal Sound ID

Personal Sound ID preserves and aligns localization cues that result in externalized sound. When sound is not externalized, it sounds similar to wearing headphones where the sound is perceived as being inside the head. Externalized sound means that the sound environment is perceived in a natural-sounding way

rather than within the listener's head. Personal Sound ID plays an important role in CrossLink Directionality 2. In listening situations that are quiet or that only consist of clear speech, CrossLink Directionality 2 will activate Personal Sound ID bilaterally to ensure the best sound quality.

P accounts for the three hearing instrument-related issues that can interfere with spatial cues that are especially salient in the high frequencies:

1. Placement of the microphones above the pinna in Behind-the-Ear (BTE) and Receiver-in-the-Ear (RIE) styles removes spectral pinna cues^{12,13}.
2. Placement of the microphones above the pinna in BTE and RIE styles distorts Interaural Level Difference (ILD)¹⁴.
3. Independently functioning Wide Dynamic Range Compression in two bilaterally fit hearing instruments can distort ILD¹⁵.

Personal Sound ID¹⁶ is modeled after the natural ear. This model includes a pinna restoration algorithm that preserves monaural spectral cues important for front-to-back and vertical localization. Increased bandwidth above 6 kHz contributes positively to localization ability¹⁷. As can be seen in Figure 2, the additional high frequency bandwidth of the Beltone Amaze compared to previous Beltone hearing aids gives access to a more complete pattern of spectral cues that emulate those of the natural pinna.

Pinna restoration high BW




Figure 2. In this chart, color represents the intensity of the sound as a function of angle of incidence (x-axis) and frequency (y-axis). The pattern of spectral cues provided by the pinna is more extensive with Beltone Amaze. The dotted line shows the upper bandwidth frequency for previous Beltone hearing aids. The cues above this cut-off would be unobtainable with earlier technology. The additional frequency bandwidth also contributes to an improvement of 1 dB in the estimation of ILD compared to previous technology.

The pinna restoration algorithm also forms the critical basis of the other component of Personal Sound ID, which is binaural compression. This part of Personal Sound ID assists in localizing sounds that occur on the left or right. This kind of localization is based in part on comparing the level differences between ears (ILD), and is also a high frequency cue. When ILD cues are consistent with interaural phase cues, externalization of sound occurs perceptually. Pinna restoration is used to estimate the ILD. Then, wireless exchange of information emulates the crossing of signals between ears, and the correction of ILD based on the ear with the least intense signal mimics inhibitory effects of auditory efferents. Due to the preserved localization cues extending to an even wider frequency bandwidth than previously in addition to other optimization of the algorithm, the average ILD error is reduced from 4.5 dB without Personal Sound ID to .7 dB with Personal Sound ID. This is a further improvement of about 1 dB compared to Personal Sound ID in Beltone Trust.

Impact on Sound Shifter

Sound Shifter uses frequency compression to improve audibility of high frequency sounds – especially speech sounds – in cases where traditional amplification is limited¹⁸. Such limitations may occur because the nature of the cochlear damage prevents the individual from being able to interpret the high frequency information. This is thought to be an issue with so-called “dead regions” of the cochlea¹⁹. Technical limitations associated with the hearing aid or fitting may also prevent sufficient amplification to make soft, high frequency speech sounds audible. Sound Shifter is a fitting option which benefits some individuals in terms of increased audibility of high frequency sounds and/or sound quality. In addition to the three settings that were available in earlier Beltone hearing aids, a “Very Mild” setting with a cut-off frequency of 5 kHz and an output bandwidth extending to 8.5 kHz widens the range of users who might experience benefit from this technology.

Benefits extend to streaming

The sound quality benefits associated with the extended bandwidth capabilities of Beltone Amaze apply to streaming of audio as well. Beltone wireless accessories have always had a wide streaming bandwidth,

with some up to 10 kHz, and with Beltone Amaze, the hearing aid can fully reproduce the bandwidth of the streamed signal. As part of the Beltone ecosystem, Beltone Amaze is compatible with existing wireless accessories. So, for example, a current Beltone First user who owns a TV Streamer 2 and upgrades to Beltone Amaze can still use their streamer with their new hearing aids. At the same time, the directly streamed sound to which the user has become accustomed will be further improved by the bandwidth capabilities of the new platform.

Like its predecessors, Beltone Amaze is also an MFi hearing aid, meaning that it is certified to work with Apple devices such as the iPhone, and can stream audio directly from these devices. As the first hearing aid manufacturer to implement this capability, Beltone has unique access to higher streaming bandwidth from these devices compared to MFi hearing aids that have since been introduced. Figure 3 shows the responses of four MFi hearing aids that were programmed as similarly as possible with flat linear gain, and presented with white noise streamed from an iPhone. Despite that the other MFi hearing instruments are capable of reproducing higher frequencies, the measurements demonstrate that the signal bandwidth is restricted. Only the Beltone Amaze (dark blue curve) shows bandwidth that corresponds to the capabilities of the hearing aid.




Figure 3. Bandwidth with direct streaming from an Apple device is limited for other MFi hearing aids. Beltone Amaze (dark blue curve) has access to a broader streaming bandwidth, contributing to sound quality.

Sound quality for new users

New users, especially those who wait to seek amplification until their hearing loss has progressed to moderate levels, have been shown to prefer less gain than those experienced hearing aid users. Further, it may take two years or more for a new user to adapt to prescribed levels of amplification²⁰. For this reason, Beltone provides fitting tools for helping those new to amplification adjust to hearing aids. The HCP can either apply a user profile that will reduce high frequency gain and increase compression ratios slightly, or simply apply a percentage of the prescribed gains to the initial fitting. Gains can gradually increase to prescribed levels automatically over a period of usage by activating the Synchronized Satisfy . Or gains can be increased manually at follow-up appointments or via Beltone Remote Care as the user becomes accustomed to the sound.

Apart from prescribed gains, those new to hearing aids may be especially sensitive to dynamic aspects of the amplification. The Beltone Curvilinear Rapid compression system uses syllabic time constants by default to increase the audibility of soft speech sounds. Although some speech sounds - such as “t” or “k” sounds - are transient in nature, the instantaneous change in energy associated with production of these sounds is limited and the Curvilinear Rapid compression system can amplify them appropriately. Other types of transient sounds have a more rapid rise and fall in sound level than speech sounds, and can be annoying for users when heard through the hearing aids. These are not the very loud “slamming door” type sounds that might trigger output limiting. And, as mentioned, they are not the softer transient speech sounds that are critical to understanding speech. Instead, sounds that are relatively loud and significantly above the background levels, like clanking dishes or jangling keys, can be of annoyance. This is due to compression time constants that may not keep up with this type of sound, resulting in a very brief period of more-than-prescribed gain when such sounds occur. Although this overshoot in amplification does not exceed uncomfortable levels, it can contribute to a perception of tinny or sharp sound quality that is a common complaint of those new to hearing aids.

Beltone Amaze introduces Impulse Noise Reduction to

prevent this issue by keeping gains at prescribed levels for brief transient sounds that might be of annoyance to users. Impulse Noise Reduction recognizes transient sounds that have rise times and energy levels which are too rapid to be speech, and is judiciously applied in such cases. Figure 4 shows the acoustic effect of Impulse Noise Reduction. Recordings were made with a Beltone Amaze programmed to a moderately severe hearing loss. The top panel shows part of the waveform for the amplified sound of plates being stacked without Impulse Noise Reduction activated. The bottom panel shows the same sound processed with Impulse Noise Reduction activated. While the reduction of the sharp, brief peaks is easy to observe on the waveform, the perceptual effect is subtle. This feature is especially useful when fitting new users because they are often fit with greater compression ratios that ensure listening comfort at louder sound levels, but that can exacerbate overshoot for impulse sounds.




Figure 4. Impulse Noise Reduction works in parallel with the Curvilinear Rapid compression system to ensure that transient sounds are not overamplified. Soft transient speech sounds are preserved. In this example, the sound of plates being stacked was recorded through a Beltone Amaze hearing aid with (bottom panel) and without (top panel) impulse noise reduction active. The light blue boxes show how peaks where overshoot occurs are reduced with impulse noise reduction.

Enhanced care

As a leader in innovation, Beltone was the first to introduce a tool that makes use of cloud-based technology that allows the HCP to provide more complete care and strengthen the relationship with the hearing aid user.

BeltoneRemoteCare²¹ is a follow-up tool that enables users to make requests or ask questions to their HCPs via the Beltone HearMax app on their smartphones. HCPs receive these requests via Beltone Solus Max software, and can respond with messages and/or fine-tuned adjustments that the user can download wirelessly to their hearing aids from their smartphones. HCPs receive requests along with information on the current hearing aid settings which can be especially informative and helpful when the user sends a request from a listening environment where they are challenged. In addition, HCPs receive a regular log of hearing aid usage from their clients' hearing aids, allowing them to proactively follow-up to encourage users who may be having issues but who may be hesitant to report them.

Summary

Although sound quality is problematic to quantify, it is highly desired by hearing aid users, and is considered a crucial attribute of hearing aids by HCPs. The Beltone philosophy for hearing aid design is to provide acoustic transparency balanced with processing for hearing loss compensation such that the listening experience is both natural and clear. With a completely new platform, Beltone Amaze leads the industry with the highest input dynamic range, and extended high frequency bandwidth. These new capabilities have comprehensive and measureable positive effects for both acoustic and streamed inputs. As a result, users can experience fuller, clearer and richer sound than ever before.

References

1. Jespersen C, Kirkwood B, Groth J. Evidence for fuller, clearer and richer sound with Beltone Amaze. Beltone white paper. 2018.
2. Kochkin S. MarkeTrak VIII: The key influencing factors in hearing aid purchase intent. Hearing Review. 2012 Mar;19(3):12-25.
3. Groth J. CrossLink Directionality 2: Directionality that supports natural auditory processing. Beltone white paper. 2016.
4. Chasin M, Russo FA. Hearing aids and music. Trends in Amplification. 2004;8(2):35-47.
5. Chasin M. What is "soft," "medium," and "loud" for speech and music: Hearing Review. 2014: February: 12. <http://www.hearingreview.com/2014/02/back-basics-soft-medium-loud-speech-music/>.
6. Moore BC, Tan CT. Perceived naturalness of spectrally distorted speech and music. J Acoust Soc Am. 2003 Jul;114(1):408-19.
7. Monson BB, Lotto AJ, Ternström S. Detection of high-frequency energy changes in sustained vowels produced by singers. J Acoust Soc Am. 2011 Apr;129(4):2263-8.
8. Best V, Carlile S, Jin C, van Schaik A. The role of high frequencies in speech localization. J Acoust Soc Am. 2005 Jul;118(1):353-63.
9. Moore BC, Füllgrabe C, Stone MA. Effect of spatial separation, extended bandwidth, and compression speed on intelligibility in a competing-speech task. J Acoust Soc Am. 2010 Jul;128(1):360-71.
10. Hayakawa S, Itakura F. The influence of noise on the speaker recognition performance using the higher frequency band. In Acoustics, Speech, and Signal Processing, 1995. ICASSP-95., 1995 International Conference on 1995 May 9 (Vol. 1, pp. 321-324). IEEE.
11. Stelmachowicz PG, Lewis DE, Choi S, Hoover B. The effect of stimulus bandwidth on auditory skills in normal-hearing and hearing-impaired children. Ear Hear. 2007 Aug;28(4):483.
12. Orton JF, Preves D. Localization as a function of hearing aid microphone placement. Hearing Instruments. 1979; 30(1); 18-21.
13. Westerman S, Topholm J. Comparing BTEs and ITEs for localizing speech. Hearing Instruments. 1985; 36(2); 20-24.
14. Udesen J, Piechowiak T, Gran F, Dittberner A. Degradation of spatial sound by the hearing aid. Proceedings of ISAAR 2013: Auditory Plasticity – Listening with the Brain. 4th symposium on audiology and Audiological Research. August 2013, Nyborg, Denmark. Dau T, Santurette S, Dalsgaard JC, Tanebjaerg L, Andersen T, Poulsen T eds.
15. Kollmeier B, Peissig J, Hovmann V. Real-time multi-band dynamic range compression and noise reduction for binaural hearing aids. Journal of Rehabilitation Research and Development. 1993; 30(1): 82-94.
16. Groth J. CrossLink Directionality with Personal Sound ID. Beltone white paper. 2014.
17. Butler RA, Planert N. The influence of stimulus bandwidth on localization of sound in space. Perception & Psychophysics. 1976 Jan 1;19(1):103-8.
18. Haastrup A. Improving high frequency audibility with Sound Shifter. Beltone white paper. 2013.
19. Moore BC. Dead regions in the cochlea: Diagnosis, perceptual consequences, and implications for the fitting of hearing aids. Trends in Amplification. 2001 Mar;5(1):1-34.
20. Keidser G, Dillon H, Carter L, O'Brien A. NAL-NL2 empirical adjustments. Trends in Amplification. 2012 Dec;16(4):211-23.
21. Stender T, Groth J, Fabry D. Teleaudiology: Friend or foe in the consumerism of hearing healthcare. Part 2: Promoting better fit to preference and efficiency. Hearing Review. 2017; 24(5). <http://www.hearingreview.com/2017/05/teleaudiology-friend-foe-consumerism-hearing-healthcare-2/>.

Worldwide headquarters

Beltone A/S
Lautrupbjerg 7
DK-2750 Ballerup, Denmark
Tel.: +45 45 75 11 11
beltone-hearing.com

CVR no. 55082715

